

NANTONG MICHAEL IRRADIATION CO., LTD


CERTIFICATE OF GAMMA STERILIZATION

Date:2022-1-15

| | |
|---|--|
| STERILE COMPANY | NANTONG MICHAEL IRRADIATION CO.,LTD |
| EXPORTER | ZHEJIANG FIRSD GROUP CO., LTD. |
| IMPORTER | Global Trade Technology – Comércio, Importação e Exportação Ltda - EPP. Av. José Antonio Miziara n 221 - Distrito Ind. José Aparecido Tomé - Jaboticabal – SP. – 14.874-002 Fone/Fax: (16) 4042-4000 CPNJ: 07.610.601/0001-65 INSCRIÇÃO ESTADUAL: 391.173.147.116 |
| 灭菌日期 DATE OF STERILIZATION | 15th Jan. 2022 |
| Invoice Number: 2022F02J04 Product: AS ATTACHMENT | |
| 要求 GAMMA 剂量 MIN GAMMA DOSE 25KGY MAXIUM GAMMA DOSE: 40KGY | |
| <p>南通市迈克尔辐照有限公司 特此证明上述产品已经本公司 GAMMA 射线消毒 NANTONG MICHAEL IRRADIATION CO., LTD hereby warranty that all products mentioned above have been sterilized by GAMMA RAY</p>  | |

STAMP:

SIGNATURE



ATTACHMENT

| CODE | PRODUCTS | QTY |
|---------|--|-----------|
| GR-01 | Swabs with polystyrene (plastic) stick, 150mm, 100% reyon head. Individual peel pack, sterile by Gamma Ray. 100pcs/bag, Lot: GR0120220113, MFG:01/2022/ DUE DATE:01/2024 | 35400BAGS |
| GT-KT06 | 2ml viral transport medium kit + nasal flocked swab and flocked swab , Lot: GTKT0620220205, MFG:02/2022/ DUE DATE:02/2024 | 20000PCS |
| GN-01 | NYLON NASAL FLOCKED SWAB; BREAK POINT DISTANCE 80MM(FROM TIP END) , Lot: GN0120220228, MFG:02/2022/ DUE DATE:02/2024 | 10000BAGS |